

See reverse side for warnings and precautions

**Is using EXTRANEAL (icodextrin 7.5%)
peritoneal dialysis solution**

Emergency Contact Information

Nephrologist	_____	()
PD Nurse/Others	_____	()
Other Contact	_____	()

BAXTER and **EXTRANEAL** are trademarks of Baxter International Inc. AL06310 12/06

WARNING AND PRECAUTION
Potential for Incorrect Blood Glucose Reading

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

DO NOT use glucose monitors or test strips that utilize the glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (icodextrin 7.5%) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events.

Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

AUGUST 2010

WC 08/10

Baxter