



Attention Hospital Physician
WARNING
Potential for Incorrect Blood Glucose Reading

July 2010

Dear Hospital Physician,

Baxter Healthcare Corporation would like to notify you of **important safety information** involving patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) who may require the use of blood glucose monitors and test strips.

Patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL) may have incorrect blood glucose results when using particular blood glucose monitors and test strips.

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

The term “glucose-specific” applies to monitors or test strips that are not affected by the presence of maltose or certain other sugars. Because Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) results in elevated blood levels of maltose, only glucose-specific monitors and test strips should be used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) due to maltose interference. A blood glucose reading with these monitors that appears to be within the normal range in a patient on Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) may mask true hypoglycemia (low blood sugar). This would cause a patient or health care professional not to take the appropriate steps to bring the blood sugar into a normal range. A falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.


Additional considerations for patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**):

1. Discontinuing Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) use will not immediately address the risk for the potential interference with glucose monitors. Falsely elevated glucose levels may result up to two weeks following cessation of Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**).
2. To determine what type of method is used for monitoring glucose levels, review the labeling for BOTH the glucose monitor and the test strips used.
If in doubt, contact the manufacturer of the glucose monitors and test strips to determine the method that is used.
3. If your hospital uses electronic medical records, the above information describing the potential for interference with blood glucose monitors or test strips needs to be entered in a suitable field readily apparent to all users.

For further information, refer to Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) prescribing information or visit www.glucosesafety.com.

I hope this information is helpful to you. If you have additional questions about Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**), please contact your Baxter Renal Representative.

Yours sincerely,



Anders Tranaeus, M.D., Ph.D
Vice President, Medical Director
Medical Affairs
Asia Pacific



Attention Hospital Nurse

WARNING

Potential for Incorrect Blood Glucose Reading

July 2010

Dear Hospital Nurse,

Baxter Healthcare Corporation would like to notify you of **important safety information** involving patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) who may require the use of blood glucose monitors and test strips.

Patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL) may have incorrect blood glucose results when using particular blood glucose monitors and test strips.

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

The term “glucose-specific” applies to monitors or test strips that are not affected by the presence of maltose or certain other sugars. Because Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) results in elevated blood levels of maltose, only glucose-specific monitors and test strips should be used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) due to maltose interference. A blood glucose reading with these monitors that appears to be within the normal range in a patient on Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) may mask true hypoglycemia (low blood sugar). This would cause a patient or health care professional not to take the appropriate steps to bring the blood sugar into a normal range. A falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.


Additional considerations for patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**):

1. Discontinuing Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) use will not immediately address the risk for the potential interference with glucose monitors. Falsely elevated glucose levels may result up to two weeks following cessation of Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**).
2. To determine what type of method is used for monitoring glucose levels, review the labeling for BOTH the glucose monitor and the test strips used.
If in doubt, contact the manufacturer of the glucose monitors and test strips to determine the method that is used.
3. If your hospital uses electronic medical records, the above information describing the potential for interference with blood glucose monitors or test strips needs to be entered in a suitable field readily apparent to all users.

For further information, refer to Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) prescribing information or visit www.glucozesafety.com.

I hope this information is helpful to you. If you have additional questions about Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**), please contact your Baxter Renal Representative.

Yours sincerely,



Anders Tranaeus, M.D., Ph.D
Vice President, Medical Director
Medical Affairs
Asia Pacific



Attention Hospital Admission
WARNING
Potential for Incorrect Blood Glucose Reading

July 2010

Dear Hospital Admission,

Baxter Healthcare Corporation would like to notify you of **important safety information** involving patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) who may require the use of blood glucose monitors and test strips.

Patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL) may have incorrect blood glucose results when using particular blood glucose monitors and test strips.

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

The term “glucose-specific” applies to monitors or test strips that are not affected by the presence of maltose or certain other sugars. Because Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) results in elevated blood levels of maltose, only glucose-specific monitors and test strips should be used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) due to maltose interference. A blood glucose reading with these monitors that appears to be within the normal range in a patient on Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) may mask true hypoglycemia (low blood sugar). This would cause a patient or health care professional not to take the appropriate steps to bring the blood sugar into a normal range. A falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.


Additional considerations for patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**):

1. Discontinuing Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) use will not immediately address the risk for the potential interference with glucose monitors. Falsely elevated glucose levels may result up to two weeks following cessation of Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**).
2. To determine what type of method is used for monitoring glucose levels, review the labeling for BOTH the glucose monitor and the test strips used.
If in doubt, contact the manufacturer of the glucose monitors and test strips to determine the method that is used.
3. If your hospital uses electronic medical records, the above information describing the potential for interference with blood glucose monitors or test strips needs to be entered in a suitable field readily apparent to all users.

For further information, refer to Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) prescribing information or visit www.glucozesafety.com.

I hope this information is helpful to you. If you have additional questions about Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**), please contact your Baxter Renal Representative.

Yours sincerely,



Anders Tranaeus, M.D., Ph.D
Vice President, Medical Director
Medical Affairs
Asia Pacific



Attention Hospital Pharmacy
WARNING
Potential for Incorrect Blood Glucose Reading

July 2010

Dear Hospital Pharmacy,

Baxter Healthcare Corporation would like to notify you of **important safety information** involving patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) who may require the use of blood glucose monitors and test strips.

Patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL) may have incorrect blood glucose results when using particular blood glucose monitors and test strips.

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

The term “glucose-specific” applies to monitors or test strips that are not affected by the presence of maltose or certain other sugars. Because Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) results in elevated blood levels of maltose, only glucose-specific monitors and test strips should be used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) due to maltose interference. A blood glucose reading with these monitors that appears to be within the normal range in a patient on Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) may mask true hypoglycemia (low blood sugar). This would cause a patient or health care professional not to take the appropriate steps to bring the blood sugar into a normal range. A falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.


Additional considerations for patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**):

1. Discontinuing Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) use will not immediately address the risk for the potential interference with glucose monitors. Falsely elevated glucose levels may result up to two weeks following cessation of Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**).
2. To determine what type of method is used for monitoring glucose levels, review the labeling for BOTH the glucose monitor and the test strips used.
If in doubt, contact the manufacturer of the glucose monitors and test strips to determine the method that is used.
3. If your hospital uses electronic medical records, the above information describing the potential for interference with blood glucose monitors or test strips needs to be entered in a suitable field readily apparent to all users.

For further information, refer to Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) prescribing information or visit www.glucozesafety.com.

I hope this information is helpful to you. If you have additional questions about Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**), please contact your Baxter Renal Representative.

Yours sincerely,



Anders Tranaeus, M.D., Ph.D
Vice President, Medical Director
Medical Affairs
Asia Pacific



Attention Laboratory Services

WARNING

Potential for Incorrect Blood Glucose Reading

July 2010

Dear Laboratory Services,

Baxter Healthcare Corporation would like to notify you of **important safety information** involving patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) who may require the use of blood glucose monitors and test strips.

Patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL) may have incorrect blood glucose results when using particular blood glucose monitors and test strips.

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

The term “glucose-specific” applies to monitors or test strips that are not affected by the presence of maltose or certain other sugars. Because Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) results in elevated blood levels of maltose, only glucose-specific monitors and test strips should be used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) due to maltose interference. A blood glucose reading with these monitors that appears to be within the normal range in a patient on Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) may mask true hypoglycemia (low blood sugar). This would cause a patient or health care professional not to take the appropriate steps to bring the blood sugar into a normal range. A falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.


Additional considerations for patients who use Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**):

1. Discontinuing Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) use will not immediately address the risk for the potential interference with glucose monitors. Falsely elevated glucose levels may result up to two weeks following cessation of Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**).
2. To determine what type of method is used for monitoring glucose levels, review the labeling for BOTH the glucose monitor and the test strips used.
If in doubt, contact the manufacturer of the glucose monitors and test strips to determine the method that is used.
3. If your hospital uses electronic medical records, the above information describing the potential for interference with blood glucose monitors or test strips needs to be entered in a suitable field readily apparent to all users.

For further information, refer to Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**) prescribing information or visit www.glucozesafety.com.

I hope this information is helpful to you. If you have additional questions about Peritoneal Dialysis Solution with 7.5% Icodextrin (**EXTRANEAL**), please contact your Baxter Renal Representative.

Yours sincerely,



Anders Tranaeus, M.D., Ph.D
Vice President, Medical Director
Medical Affairs
Asia Pacific