

## Country-Specific Glucose Monitor List

**Country Name:** *Puerto Rico*

### **Important Information**

This is a non-comprehensive list, current as of January 2017. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with icodextrin and maltose, Baxter does NOT recommend that EXTRANEAL patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitor manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with maltose and icodextrin to Baxter's recommended limits: 278 mg/dL (maltose) and 1094 mg/dL (icodextrin). The manufacturers certified that their "green" monitors, below, showed no interference of blood glucose readings under these conditions, with the exception of the "green" monitors specifically notated <sup>4</sup> for which certification of testing to Baxter's recommended limits has not yet been received. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) peritoneal dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains icodextrin. Maltose, a metabolite of icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Or, a falsely-elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by icodextrin or maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. **ONLY glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution.** Contact the manufacturer to verify that the test strips and monitor are glucose-specific. The list provides the contact information of the more common, major brand manufacturers.

## GLUCOSE MONITORS

Updated March 2016

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose- specific)	Test Type*	Manufacturer
Boots <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	Abbott Diabetes Care <a href="http://www.abbottdiabetescare.com">www.abbottdiabetescare.com</a> Phone: 888-522-5226
FreeStyle Flash	Yes	GDH-FAD	
FreeStyle Freedom	Yes	GDH-FAD	
FreeStyle Freedom Lite	Yes	GDH-FAD	
FreeStyle InsuLinx	Yes	GDH-FAD	
FreeStyle Lite	Yes	GDH-FAD	
FreeStyle Mini	Yes	GDH-FAD	
FreeStyle Optium <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
FreeStyle Optium H <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
FreeStyle Optium Neo <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
FreeStyle Optium Neo H <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
FreeStyle Papillon InsuLinx <sup>5</sup>	Yes <sup>5</sup>	GDH-FAD <sup>5</sup>	
FreeStyle Papillon Lite <sup>5</sup>	Yes <sup>5</sup>	GDH-FAD <sup>5</sup>	
FreeStyle Papillon Mini <sup>5</sup>	Yes <sup>5</sup>	GDH-FAD <sup>5</sup>	
FreeStyle Papillon Vision <sup>5</sup>	Yes <sup>5</sup>	GDH-FAD <sup>5</sup>	
FreeStyle Precision <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
FreeStyle Precision H	Yes	GDH-NAD	
FreeStyle Precision Neo	Yes	GDH-NAD	
FreeStyle Precision Pro	Yes	GDH-NAD	
Omron HEA-214 <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Optium <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Optium Easy <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
OptiumEZ	Yes	GDH-NAD	
Optium Xceed <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Optium Xido <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Optium Xido Neo <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Precision Xceed <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Precision Xceed Pro <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Precision Xtra	Yes	GDH-NAD	
Precision Xtra OK <sup>1, 5</sup>	Yes <sup>1, 5</sup>	GDH-NAD, GO <sup>1, 5</sup>	
ReliOn Ultima	Yes	GDH-NAD	
TrueSense <sup>5</sup>	Yes <sup>5</sup>	GDH-NAD <sup>5</sup>	
Assure Platinum	Yes	GO	Arkray, Inc. <a href="http://www.arkrayusa.com">www.arkrayusa.com</a> Phone: 800-818-8877
Assure Prism	Yes	GO	
GLUCOCARD 01	Yes	GO	
GLUCOCARD 01-mini	Yes	GO	
GLUCOCARD 01-mini plus	Yes	GO	
GLUCOCARD Expression	Yes	GO	
GLUCOCARD G Black	Yes	GDH-FAD	
GLUCOCARD MX	Yes	GDH-FAD	
GLUCOCARD MyDIA	Yes	GO	
GLUCOCARD Prism <sup>4</sup>	Yes <sup>4</sup>	GO <sup>4</sup>	
GLUCOCARD Shine	Yes	GO	
GLUCOCARD S	Yes	GDH-FAD	
GLUCOCARD SM	Yes	GDH-FAD	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
<b>GLUCOCARD Vital</b> <b>GLUCOCARD X-meter<sup>2</sup></b> <b>GLUCOCARD X-mini<sup>2</sup></b> <b>GLUCOCARD X-mini plus<sup>2</sup></b> <b>GLUCOCARD Σ</b> <b>GLUCOCARD Σ-mini</b> <b>ReliOn Confirm</b> <b>ReliOn micro</b> <b>ReliOn Prime</b>	<b>Yes</b> <b>Yes<sup>2</sup></b> <b>Yes<sup>2</sup></b> <b>Yes<sup>2</sup></b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GDH-FAD<sup>2</sup></b> <b>GDH-FAD<sup>2</sup></b> <b>GDH-FAD<sup>2</sup></b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b>	
<b>Ascensia Brio</b> <b>Ascensia Entrust</b> <b>Breeze 2</b> <b>Contour</b> <b>Contour Link</b> <b>Contour Next</b> <b>Contour Next EZ</b> <b>Contour Next Link</b> <b>Contour Next Link 2.4</b> <b>Contour Next USB</b> <b>Contour Plus</b> <b>Contour Plus Link 2.4</b> <b>Contour TS</b> <b>Contour USB</b> <b>Contour XT</b> <b>Elite<sup>A</sup></b> <b>Elite XL</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GO</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GO</b> <b>GO</b>	<b>Ascensia Diabetes Care</b> (formerly Bayer Healthcare) <a href="http://www.ascensia.com">www.ascensia.com</a> <b>Phone: 800-348-8100</b>
<b>OneTouch InDuo</b> <b>OneTouch Ping</b> <b>OneTouch Select</b> <b>OneTouch Select Mini</b> <b>OneTouch Select Simple</b> <b>OneTouch SureStep</b> <b>OneTouch Ultra<sup>B</sup></b> <b>OneTouch Ultra 2</b> <b>OneTouch UltraEasy</b> <b>OneTouch UltraLink</b> <b>OneTouch UltraMini</b> <b>OneTouch SelectPlus</b> <b>OneTouch SelectPlus Flex</b> <b>OneTouch UltraSmart</b> <b>OneTouch UltraVue</b> <b>OneTouch Verio</b> <b>OneTouch VerioFlex</b> <b>OneTouch VerioIQ</b> <b>OneTouch VerioPro</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Lifescan, Inc.</b> <a href="http://www.lifescan.com">www.lifescan.com</a> <b>Phone: 800-227-8862</b>

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
OneTouch VerioPro+ OneTouch VerioSync OneTouch VerioVue OneTouch Vita SureStep Flexx	Yes Yes Yes Yes Yes	GDH-FAD GDH-FAD GDH-FAD GO GO	
Nova Max Plus Nova Max Link StatStrip Hospital Glucose Meter StatStrip Hospital Glucose and Ketone Monitoring System <sup>5</sup> StatStrip Xpress Glucose Meter StatStrip Xpress Glucose and Ketone Monitoring System <sup>5</sup> StatStrip Xpress2 Glucose Meter Nova Pro Glucose/Ketone Meter <sup>5</sup>	Yes Yes Yes Yes <sup>5</sup> Yes Yes <sup>5</sup> Yes Yes <sup>5</sup>	GO GO GO GO <sup>5</sup> GO GO <sup>5</sup> GO GDH-FAD <sup>5</sup>	Nova Biomedical <a href="http://www.novabiomedical.com">www.novabiomedical.com</a> Phone: (781) 894-0800
Accu-Chek Active Accu-Chek Aviva Accu-Chek Aviva Combo Accu-Chek Aviva Connect Accu-Chek Aviva Expert Accu-Chek Aviva Insight Accu-Chek Aviva Nano Accu-Chek Aviva Plus <sup>3</sup> Accu-Chek Compact Plus Accu-Chek Guide Accu-Chek Inform II Accu-Chek Nano <sup>3</sup> Accu-Chek Nano SmartView <sup>3</sup> Accu-Chek Mobile Accu-Chek Performa Accu-Chek Performa Combo Accu-Chek Performa Connect Accu-Chek Performa Insight Accu-Chek Performa Nano Accu-Chek Voicemate Plus System	Yes Yes Yes Yes Yes Yes Yes Yes <sup>3</sup> Yes Yes Yes Yes <sup>3</sup> Yes <sup>3</sup> Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut-Q-GDH <sup>3</sup> Mut Q-GDH GDH-FAD Mut Q-GDH Mut Q-GDH <sup>3</sup> Mut Q-GDH <sup>3</sup> Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH	Roche Diagnostics <a href="http://www.accu-chek.com">www.accu-chek.com</a> Phone: 800-858-8072
Accu-Chek Go/Go S System Accu-Chek Integra System	No No	GDH-PQQ GDH-PQQ	
Jazz Jazz Wireless Presto iBGStar	Yes Yes Yes Yes	GO GO GO GO	AgaMatrix, Inc. <a href="http://www.agamatrix.com">www.agamatrix.com</a> Phone : 866-906-4197
TRUE Metrix <sup>4</sup> TRUE Metrix Air <sup>4</sup> TRUE Track <sup>4</sup>	Yes <sup>4</sup> Yes <sup>4</sup> Yes <sup>4</sup>	GDH-FAD <sup>4</sup> GDH-FAD <sup>4</sup> GDH-FAD <sup>4</sup> GDH-FAD <sup>4</sup>	Trividia Health (formerly Nipro Diagnostics) <a href="http://www.trividiahealth.com">http://www.trividiahealth.com</a> om/Phone: 800-803-6025

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
TRUE Balance <sup>4</sup>	Yes <sup>4</sup>		
Prodigy AutoCode <sup>4</sup> Prodigy Voice <sup>4</sup> Prodigy Pocket <sup>4</sup> Prodigy No Coding Test Strips <sup>4</sup>	Yes <sup>4</sup> Yes <sup>4</sup> Yes <sup>4</sup> Yes <sup>4</sup>	GO <sup>4</sup> GO <sup>4</sup> GO <sup>4</sup> GO <sup>4</sup>	Prodigy Diabetes Care <a href="mailto:customercare@prodigymeter.com">customercare@prodigymeter.com</a> Phone: 800-243-2636

<sup>1</sup> Two types of **compatible** test strips for Precision Xtra OK.

<sup>2</sup> These Arkray monitors/test strips have switched from an incompatible GDH-FAD based chemistry to a **compatible** GDH-FAD based chemistry. Consult manufacturer for additional information.

<sup>3</sup> The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY, and use test strips branded as ACCU-CHEK Smartview and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**compatible**) strips. Consult manufacturer for additional information.

<sup>4</sup> This/These monitor/test strips were not currently certified as having been tested to Baxter's recommended interference limits for maltose or icodextrin when this List was issued. Consult manufacturer for additional information

<sup>5</sup> These brand name monitors are not available within United States.

#### Test Type\*

**GO** = glucose oxidase

**GDH-PQQ** = glucose dehydrogenase with pyrroloquinolinequinone (note: **GDO**, glucose-dye-oxidoreductase, is an **incompatible** PQQ-based method)

**GDH-NAD** = glucose dehydrogenase with nicotinamide-adenine dinucleotide

**GDH-FAD** = glucose dehydrogenase with flavin-adenine dinucleotide

**Mut Q-GDH** = glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

#### References:

<sup>A</sup> Baxter report REP-NIV-RE-366 Evaluation of potential interference in blood glucose determination (measured with enzymatic methods) for patients treated with icodextrin.

<sup>B</sup> Baxter report Interim 1, 33541 Determination of potential interference of icodextrin and its metabolites on human blood glucose measurement using chosen glucometers.

Please see full prescribing information for EXTRANEAL (icodextrin) PD solution.

Baxter and EXTRANEAL are trademarks of Baxter International Inc.

Other trademarks or products appearing herein are the property of their respective owners.