

Country-Specific Glucose Monitor List

Country Name: *United States*

Important Information

This is a non-comprehensive list, current as of *June 2017*. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (icodextrin) Peritoneal Dialysis Solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL (icodextrin) Solution. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with icodextrin and maltose, Baxter does NOT recommend that EXTRANEAL (icodextrin) Solution patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitor manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with maltose and icodextrin to Baxter's recommended limits: 278 mg/dL (maltose) and 1094 mg/dL (icodextrin). The manufacturers certified that their "green" monitors, below, showed no interference of blood glucose readings under these conditions, with the exception of the "green" monitors specifically notated⁴ for which certification of testing to Baxter's recommended limits has not yet been received. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacturer's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (icodextrin) PD Solution.

1. EXTRANEAL (icodextrin) solution contains icodextrin. Maltose, a metabolite of icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (icodextrin) solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely-elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by icodextrin or maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. **ONLY glucose-specific monitors and test strips should be used with patients on EXTRANEAL (icodextrin) solution.** Contact the manufacturer to verify that the test strips and monitor are glucose-specific. The list provides the contact information of the more common, major brand manufacturers.

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

GLUCOSE MONITORS

Updated June 2017

Glucose Monitor Brand	Compatible with EXTRANEAL (icodextrin) PD Solution (Glucose- specific)	Test Type*	Manufacturer
Boots ⁵ FreeStyle Flash FreeStyle Freedom FreeStyle Freedom Lite FreeStyle InsuLinx FreeStyle Lite FreeStyle Mini FreeStyle Optium ⁵ FreeStyle Optium H ⁵ FreeStyle Optium Neo ⁵ FreeStyle Optium Neo H ⁵ FreeStyle Papillon InsuLinx ⁵ FreeStyle Papillon Lite ⁵ FreeStyle Papillon Mini ⁵ FreeStyle Papillon Vision ⁵ FreeStyle Precision ⁵ FreeStyle Precision H FreeStyle Precision Neo FreeStyle Precision Pro Omron HEA-214 ⁵ Optium ⁵ Optium Easy ⁵ OptiumEZ Optium Xceed ⁵ Optium Xido ⁵ Optium Xido Neo ⁵ Precision Xceed ⁵ Precision Xceed Pro ⁵ Precision Xtra Precision Xtra OK ^{1,5} ReliOn Ultima TrueSense ⁵	Yes ⁵ Yes Yes Yes Yes Yes Yes Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes Yes Yes Yes Yes ⁵ Yes ⁵ Yes ⁵ Yes Yes Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes ⁵ Yes Yes ^{1,5} Yes Yes ⁵	GDH-NAD ⁵ GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-FAD ⁵ GDH-FAD ⁵ GDH-FAD ⁵ GDH-FAD ⁵ GDH-NAD ⁵ GDH-NAD GDH-NAD GDH-NAD GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD ⁵ GDH-NAD GDH-NAD, GO ^{1,5} GDH-NAD GDH-NAD ⁵	Abbott Diabetes Care www.abbottdiabetescare.com Phone: 888-522-5226
Assure Platinum Assure Prism ⁴ GLUCOCARD 01 GLUCOCARD 01-mini GLUCOCARD 01-mini plus GLUCOCARD Expression GLUCOCARD G Black GLUCOCARD MX GLUCOCARD MyDIA GLUCOCARD Shine ⁴	Yes Yes ⁴ Yes Yes Yes Yes Yes Yes Yes Yes ⁴	GO GO ⁴ GO GO GO GO GDH-FAD GDH-FAD GO GO ⁴	Arkray, Inc. www.arkrayusa.com Phone: 800-818-8877

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

Glucose Monitor Brand	Compatible with EXTRANEAL (icodextrin) PD Solution (Glucose-specific)	Test Type*	Manufacturer
GLUCOCARD S GLUCOCARD SM GLUCOCARD Vital GLUCOCARD X-meter² GLUCOCARD X-mini² GLUCOCARD X-mini plus² GLUCOCARD Σ GLUCOCARD Σ-mini ReliOn Confirm ReliOn micro ReliOn Prime	Yes Yes Yes Yes² Yes² Yes² Yes Yes Yes Yes Yes	GDH-FAD GDH-FAD GO GDH-FAD² GDH-FAD² GDH-FAD² GO GO GO GO GO	
Breeze 2 Contour Contour Link Contour Next Contour Next EZ Contour Next Link Contour Next Link 2.4 Contour Next USB Contour Plus Contour Plus Link 2.4 Contour TS Contour USB Contour XT Contour Next One Blood Contour Plus One Blood	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GO GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Ascensia Diabetes Care (formerly Bayer Healthcare) www.ascensia.com Phone: 800-348-8100
OneTouch InDuo OneTouch Select OneTouch Select Mini OneTouch Select Simple OneTouch Ultra^A OneTouch Ultra 2 OneTouch UltraEasy OneTouch UltraLink OneTouch UltraMini OneTouch SelectPlus OneTouch SelectPlus Flex OneTouch UltraSmart OneTouch UltraVue OneTouch Verio OneTouch VerioFlex OneTouch VerioIQ OneTouch VerioPro	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GO GO GO GO GO GO GO GO GO GO GO GO GO GO GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Lifescan, Inc. www.lifescan.com Phone: 800-227-8862

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

Glucose Monitor Brand	Compatible with EXTRANEAL (icodextrin) PD Solution (Glucose-specific)	Test Type*	Manufacturer
OneTouch VerioPro+ OneTouch VerioSync OneTouch VerioVue OneTouch Vita	Yes Yes Yes Yes	GDH-FAD GDH-FAD GDH-FAD GO	
Nova Max Plus Nova Max Link StatStrip Hospital Glucose Meter StatStrip Hospital Glucose and Ketone Monitoring System ⁵ StatStrip Xpress Glucose Meter StatStrip Xpress Glucose and Ketone Monitoring System ⁵ StatStrip Xpress2 Glucose and Ketone Monitoring Meter Nova Pro Glucose/Ketone Meter ⁵	Yes Yes Yes Yes ⁵ Yes Yes ⁵ Yes Yes ⁵	GO GO GO GO ⁵ GO GO ⁵ GO GDH-FAD ⁵	Nova Biomedical www.novabiomedical.com Phone: (781) 894-0800
Accu-Chek Active Accu-Chek Aviva Accu-Chek Aviva Combo Accu-Chek Aviva Connect Accu-Chek Aviva Expert Accu-Chek Aviva Insight Accu-Chek Aviva Nano Accu-Chek Aviva Plus ³ Accu-Chek Compact Plus Accu-Chek Guide Accu-Chek Inform II Accu-Chek Instant Accu-Chek Instant S Accu-Chek Nano ³ Accu-Chek Nano SmartView ³ Accu-Chek Mobile Accu-Chek Performa Accu-Chek Performa Combo Accu-Chek Performa Connect Accu-Chek Performa Insight Accu-Chek Performa Nano Accu-Chek Voicemate Plus System	Yes Yes Yes Yes Yes Yes Yes Yes ³ Yes Yes Yes Yes Yes Yes ³ Yes ³ Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut-Q-GDH ³ Mut Q-GDH GDH-FAD Mut Q-GDH GDH-FAD GDH-FAD Mut Q-GDH ³ Mut Q-GDH ³ Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH	Roche Diagnostics www.accu-chek.com Phone: 800-858-8072
Jazz Jazz Wireless 2 Presto iBGStar One Drop	Yes Yes Yes Yes Yes	GO GO GO GO GO	AgaMatrix, Inc. www.agamatrix.com Phone : 866-906-4197

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

Glucose Monitor Brand	Compatible with EXTRANEAL (icodextrin) PD Solution (Glucose-specific)	Test Type*	Manufacturer
TRUE Metrix ⁴ TRUE Metrix Air ⁴ TRUE Track ⁴ TRUE Balance ⁴	Yes ⁴ Yes ⁴ Yes ⁴ Yes ⁴	GDH-FAD ⁴ GDH-FAD ⁴ GDH-FAD ⁴ GDH-FAD ⁴	Trividia Health (formerly Nipro Diagnostics) www.trividiahealth.com Phone: 800-803-6025
Prodigy AutoCode ⁴ Prodigy Voice ⁴ Prodigy Pocket ⁴ Prodigy No Coding Test Strips ⁴	Yes ⁴ Yes ⁴ Yes ⁴ Yes ⁴	GO ⁴ GO ⁴ GO ⁴ GO ⁴	Prodigy Diabetes Care www.prodigymeter.com Phone: 800-243-2636

¹ Two types of **compatible** test strips for Precision Xtra OK.

² These Arkray monitors/test strips have switched from an **incompatible** GDH-FAD based chemistry to a **compatible** GDH-FAD based chemistry. Consult manufacturer for additional information.

³ The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY, and use test strips branded as ACCU-CHEK Smartview and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**compatible**) strips. Consult manufacturer for additional information.

⁴ This/These monitor/test strips were not currently certified as having been tested to Baxter's recommended interference limits for maltose or icodextrin when this List was issued. Consult manufacturer for additional information

⁵ These brand name monitors are not available within United States.

Test Type*

GO = glucose oxidase

GDH-PQQ = glucose dehydrogenase with pyrroloquinolinequinone (note: **GDO**, glucose-dye-oxidoreductase, is an **incompatible** PQQ-based method)

GDH-NAD = glucose dehydrogenase with nicotinamide-adenine dinucleotide

GDH-FAD = glucose dehydrogenase with flavin-adenine dinucleotide

Mut Q-GDH = glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

References:

^A Baxter report Interim 1, 33541 Determination of potential interference of icodextrin and its metabolites on human blood glucose measurement using chosen glucometers.

Please see full prescribing information for EXTRANEAL (icodextrin) PD solution.

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

EXTRANEAL (icodextrin) Peritoneal Dialysis Solution Indication and Important Risk Information (IRI)

Indication:

EXTRANEAL (icodextrin) is indicated for a single daily exchange for the long (8- to 16- hour) dwell during continuous ambulatory peritoneal dialysis (CAPD) or automated peritoneal dialysis (APD) for the management of end-stage renal disease. EXTRANEAL (icodextrin) is also indicated to improve (compared to 4.25% dextrose) long-dwell ultrafiltration and clearance of creatinine and urea nitrogen in patients with high average or greater transport characteristics, as defined using the peritoneal equilibration test (PET).

Important Risk Information:

WARNING: UNRECOGNIZED HYPOGLYCEMIA RESULTING FROM DRUG-DEVICE INTERACTION

- **Only use glucose-specific monitors and test strips to measure blood glucose levels in patients using EXTRANEAL (icodextrin) Peritoneal Dialysis Solution. Blood glucose monitoring devices using glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase (GDO)-based methods must not be used. In addition, some blood glucose monitoring systems using glucose dehydrogenase flavin-adenine dinucleotide (GDHFAD)-based methods must not be used. Use of GDH-PQQ, GDO, and GDH-FAD-based glucose monitors and test strips has resulted in falsely elevated glucose readings (due to the presence of maltose). Falsely elevated glucose readings have led patients or health care providers to withhold treatment of hypoglycemia or to administer insulin inappropriately. Both of these situations have resulted in unrecognized hypoglycemia, which has led to loss of consciousness, coma, permanent neurological damage, and death. Plasma levels of EXTRANEAL (icodextrin) and its metabolites return to baseline within approximately 14 days following cessation of EXTRANEAL (icodextrin) administration. Therefore falsely elevated glucose levels may be measured up to two weeks following cessation of EXTRANEAL (icodextrin) therapy when GDH-PQQ, GDO, and GDH-FAD-based blood glucose monitors and test strips are used.**
- **To avoid improper insulin administration, educate all patients to alert health care providers of this interaction particularly in hospital settings.**
- **The manufacturer(s) of the monitor and test strips should be contacted to determine if icodextrin or maltose causes interference or falsely elevated glucose readings. For a list of toll free numbers for glucose monitor and test strip manufacturers, please contact the Baxter Renal Clinical Help Line 1-888-RENAL-HELP or visit www.glucosesafety.com.**
- **Because of the risk of unrecognized hypoglycemia that could result from a drug-device interaction, EXTRANEAL (icodextrin) is available only through a restricted program.**

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).

- EXTRANEAL (icodextrin) is contraindicated in patients with a known allergy to cornstarch or icodextrin, in patients with maltose or isomaltose intolerance, in patients with glycogen storage disease, and in patients with severe lactic acidosis.
- EXTRANEAL is intended for intraperitoneal administration only. Not for intravenous injection. Aseptic technique should be used throughout the peritoneal dialysis procedure.
- Encapsulating peritoneal sclerosis (EPS), sometimes fatal, is a complication of peritoneal dialysis therapy and has been reported in patients using EXTRANEAL.
- Serious hypersensitivity reactions to EXTRANEAL have been reported such as toxic epidermal necrolysis, angioedema, serum sickness, erythema multiforme and vasculitis. Anaphylactic or anaphylactoid reactions may occur. If a serious reaction is suspected, discontinue EXTRANEAL immediately and institute appropriate therapeutic countermeasures.
- Overinfusion of peritoneal dialysis solution volume into the peritoneal cavity may be characterized by abdominal distention, feeling of fullness and/or shortness of breath. Drain the peritoneal dialysis solution from the peritoneal cavity to treat overinfusion.
- Patients with insulin-dependent diabetes may require modification of insulin dosage following initiation of treatment with EXTRANEAL. Monitor blood glucose and adjust insulin, if needed.
- Peritoneal dialysis may affect a patient's protein, water-soluble vitamin, potassium, sodium, chloride, bicarbonate, and magnesium levels and volume status. Monitor electrolytes and blood chemistry periodically. Monitor fluid status to avoid hyper- or hypovolemia and potentially severe consequences including congestive heart failure, volume depletion, and hypovolemic shock. Abnormalities in any of these parameters should be treated promptly under the care of a physician.
- In clinical trials, the most frequently reported adverse events occurring in $\geq 10\%$ of patients and more common in EXTRANEAL PD Solution patients than in control patients, were peritonitis, upper respiratory infection, hypertension, and rash. The most common treatment-related adverse reaction for EXTRANEAL PD Solution patients was skin rash.

Please see full Prescribing Information at www.baxter.com.

Baxter and Extraneal are trademarks of Baxter International Inc.

Any other trademarks or products appearing herein are the property of their respective owners.

Please see Indication and Important Risk Information for EXTRANEAL (icodextrin) PD Solution, including BOXED WARNING on pages 6-7. Please see the full [Prescribing Information](#) and [Medication Guide](#).