

Country-Specific Glucose Monitor List

Country Name: *India*

Important Information

This is a non-comprehensive list, current as of **14th January 2020**. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with icodextrin and maltose, Baxter does NOT recommend that EXTRANEAL patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitor manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with maltose and icodextrin to Baxter's recommended limits: 278 mg/dL (maltose) and 1094 mg/dL (icodextrin). The manufacturers certified that their "green" monitors, below, showed no interference of blood glucose readings under these conditions. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) peritoneal dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains icodextrin. Maltose, a metabolite of icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely-elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by icodextrin or maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. **ONLY glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution.** Contact the manufacturer to verify that the test strips and monitor are glucose-specific. The list provides the contact information of the more common, major brand manufacturers.

GLUCOSE MONITORS

Updated 14th January 2020

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose- specific)	Test Type*	Manufacturer
FreeStyle Freedom	Yes	GDH-FAD	Abbott Diabetes Care http://in.abbott-diabetescare.com/ Phone: 0008001005780 Address: Abbott Healthcare Pvt. Ltd. Abbott Diabetes Care 3-4 Corporate Park Sion-Trombay Road Mumbai, Pin-400071 India
FreeStyle Freedom Lite	Yes	GDH-FAD	
FreeStyle Lite	Yes	GDH-FAD	
FreeStyle Libre ¹	Not recommended	GO	
FreeStyle Libre Pro ¹	Not recommended	GO	
FreeStyle Optium Neo	Yes	GDH-NAD	
FreeStyle Optium Neo H	Yes	GDH-NAD	
FreeStyle Papillon Vision	Yes	GDH-FAD	
FreeStyle Precision Neo	Yes	GDH-NAD	
FreeStyle Precision Pro	Yes	GDH-NAD	
Optium Xido Neo	Yes	GDH-NAD	
Precision Xceed Pro	Yes	GDH-NAD	
Assure Platinum	Yes	GO	Arkray, Inc. http://www.arkray.co.in/ Phone: +912261559355 Address: 701 / 702, Opulence, TPS III, 6th Road, Santacruz East, Mumbai, Maharashtra 400055, India
GLUCOCARD 01	Yes	GO	
GLUCOCARD 01-mini	Yes	GO	
GLUCOCARD 01-mini plus	Yes	GO	
GLUCOCARD G Black	Yes	GDH-FAD	
GLUCOCARD MX	Yes	GDH-FAD	
GLUCOCARD MyDIA	Yes	GO	
GLUCOCARD S	Yes	GDH-FAD	
GLUCOCARD SM	Yes	GDH-FAD	
GLUCOCARD Vital	Yes	GO	
GLUCOCARD X-meter ²	Yes ²	GDH-FAD ²	
GLUCOCARD X-mini ²	Yes ²	GDH-FAD ²	
GLUCOCARD X-mini plus ²	Yes ²	GDH-FAD ²	
GLUCOCARD Σ	Yes	GO	
GLUCOCARD Σ-mini	Yes	GO	
ReliOn Confirm	Yes	GO	
ReliOn micro	Yes	GO	
ReliOn Prime	Yes	GO	
Breeze 2	Yes	GO	Ascensia Diabetes Care (formerly Bayer Healthcare) https://www.diabetes.ascensia.in/ Email: dcindia@ascensia.com Phone: 1800 123 0123 Address: Ascensia Diabetes Care India Private Limited 1st Floor, Bella Vista, Oswal Park Pokhran Road No-2, Majiwada Junction Thane (W) – 400 601 Maharashtra, India
Contour	Yes	GDH-FAD	
Contour Link	Yes	GDH-FAD	
Contour Next	Yes	GDH-FAD	
Contour Next EZ	Yes	GDH-FAD	
Contour Next Link	Yes	GDH-FAD	
Contour Next Link 2.4	Yes	GDH-FAD	
Contour Next USB	Yes	GDH-FAD	
Contour Next One Blood	Yes	GDH-FAD	
Contour Plus	Yes	GDH-FAD	
Contour Plus Link 2.4	Yes	GDH-FAD	
Contour Plus One Blood	Yes	GDH-FAD	
Contour TS	Yes	GDH-FAD	
Contour USB	Yes	GDH-FAD	
Contour XT	Yes	GDH-FAD	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose- specific)	Test Type*	Manufacturer
OneTouch Select	Yes	GO	Lifescan, Inc. http://www.onetouch.in/ Phone: 1800 22 55 44 email id: service.india@Lifescan.com Address: Unit 1003, A Wing, 10th floor, Kanakia Wall Street, Andheri East, Andheri Kurla Road , Mumbai 400093
OneTouch Select Mini	Yes	GO	
OneTouch SelectPlus	Yes	GO	
OneTouch SelectPlus Flex	Yes	GO	
OneTouch SelectPlus Simple	Yes	GO	
OneTouch Select Simple	Yes	GO	
OneTouch Ultra ^A	Yes	GO	
OneTouch Ultra 2	Yes	GO	
OneTouch UltraEasy	Yes	GO	
OneTouch UltraLink	Yes	GO	
OneTouch UltraMini	Yes	GO	
OneTouch UltraPlus Flex	Yes	GDH-FAD	
OneTouch UltraPlus Reflect	Yes	GDH-FAD	
OneTouch UltraSmart	Yes	GO	
OneTouch UltraVue	Yes	GO	
OneTouch Verio	Yes	GDH-FAD	
OneTouch VerioFlex	Yes	GDH-FAD	
OneTouch VerioIQ	Yes	GDH-FAD	
OneTouch VerioPro	Yes	GDH-FAD	
OneTouch VerioPro+	Yes	GDH-FAD	
OneTouch VerioSync	Yes	GDH-FAD	
OneTouch VerioVue	Yes	GDH-FAD	
OneTouch Verio Reflect	Yes	GDH-FAD	
Nova Max Plus	Yes	GO	Nova Biomedical www.novabiomedical.com Phone : (022) 25445000 Address: 307 Apra Plaza II Plot 14 Sector 10 Dwarka New Delhi, 110 075 India
Nova Max Link	Yes	GO	
StatStrip Hospital Glucose Meter	Yes	GO	
StatStrip Hospital Glucose and Ketone Monitoring System	Yes	GO	
StatStrip Xpress Glucose Meter	Yes	GO	
StatStrip Xpress Glucose and Ketone Monitoring System	Yes	GO	
StatStrip Xpress2 Glucose and Ketone Monitoring System	Yes	GO	
Nova Pro Glucose/Ketone Meter	Yes	GDH-FAD	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Accu-Chek Active	Yes	Mut Q-GDH	Roche Diagnostics www.roche-diagnostics.co.in Phone: +91 (22) 66974900 Address: Roche Diagnostics India Pvt Ltd 501 B, Silver Utopia, Cardinal Gracious Road, Chakala Road, Andheri East, Mumbai – 400069
Accu-Chek Aviva	Yes	Mut Q-GDH	
Accu-Chek Aviva Combo	Yes	Mut Q-GDH	
Accu-Chek Aviva Connect	Yes	Mut Q-GDH	
Accu-Chek Aviva Expert	Yes	Mut Q-GDH	
Accu-Chek Aviva Insight	Yes	Mut Q-GDH	
Accu-Chek Aviva Nano	Yes	Mut Q-GDH	
Accu-Chek Aviva Plus ³	Yes ³	Mut-Q-GDH ³	
Accu-Chek Aviva Solo	Yes	Mut Q-GDH	
Accu-Chek Compact Plus	Yes	Mut Q-GDH	
Accu-Chek Guide	Yes	GDH-FAD	
Accu-Chek Guide Me	Yes	GDH-FAD	
Accu-Chek Inform II	Yes	Mut Q-GDH	
Accu-Chek Instant	Yes	GDH-FAD	
Accu-Chek Instant S	Yes	GDH-FAD	
Accu-Chek Nano ³	Yes ³	Mut Q-GDH ³	
Accu-Chek Nano SmartView ³	Yes ³	Mut Q-GDH ³	
Accu-Chek Mobile	Yes	Mut Q-GDH	
Accu-Chek Performa	Yes	Mut Q-GDH	
Accu-Chek Performa Combo	Yes	Mut Q-GDH	
Accu-Chek Performa Connect	Yes	Mut Q-GDH	
Accu-Chek Performa Insight	Yes	Mut Q-GDH	
Accu-Chek Performa Nano	Yes	Mut Q-GDH	
Accu-Chek Performa Solo	Yes	Mut Q-GDH	
Accu-Chek Voicemate Plus System	Yes	Mut Q-GDH	

¹These Abbott monitors (together with the Dexcom G4 Platinum, G5 Mobile, G6 and the Medtronic iPro2) are the leading Continuous Glucose Monitor Systems (CGM) and are **not recommended** by their manufacturers for use into dialysis patients, and despite having a GO based chemistry currently there is not definitive scientific evidence for their Extraneal compatibility. Consult manufacturer for additional information.

²These Arkray monitors/test strips have switched from an **incompatible** GDH-FAD based chemistry to a **compatible** GDH-FAD based chemistry. Consult manufacturer for additional information.

³The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY and use test strips branded as ACCU-CHEK Smartview and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**compatible**) strips. Consult manufacturer for additional information.

Test Type*

GO = glucose oxidase

GDH-PQQ = glucose dehydrogenase with pyrroloquinolinequinone (note: **GDO**, glucose-dye-oxidoreductase, is an **incompatible** PQQ-based method)

GDH-NAD = glucose dehydrogenase with nicotinamide-adenine dinucleotide

GDH-FAD = glucose dehydrogenase with flavin-adenine dinucleotide

Mut Q-GDH = glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

References:

^A Baxter report Interim 1, 33541 Determination of potential interference of icodextrin and its metabolites on human blood glucose measurement using chosen glucometers.

Please see full prescribing information for EXTRANEAL (icodextrin) PD solution.

Baxter and EXTRANEAL are trademarks of Baxter International Inc.

Other trademarks or products appearing herein are the property of their respective owners.