

## WARNING

### Potential for Incorrect Blood Glucose Reading

**ONLY** use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

**DO NOT** use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (icodextrin) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events

Visit [www.glucosafety.com](http://www.glucosafety.com) for additional information, including a glucose monitor compatibility list or call the Toll Free No. 1800-102-1110."

**Vantive**

See reverse side for warnings

Patient Name

is using **EXTRANEAL** (icodextrin)  
peritoneal dialysis solution

Emergency Contact Information

Dialysis Unit: \_\_\_\_\_ ( )

Other Contact: \_\_\_\_\_ ( )

Vantive and Extraneal are trademarks of Vantive Health LLC or its affiliates.  
Date of Preparation Aug 2024

## WARNING

### Potential for Incorrect Blood Glucose Reading

is using **EXTRANEAL** (icodextrin)  
peritoneal dialysis solution

**ONLY** use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

**DO NOT** use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (icodextrin) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events.

Visit [www.glucosafety.com](http://www.glucosafety.com) for additional information, including a glucose monitor compatibility list or call the Toll Free No. 1800-102-1110."

Date of Preparation Aug 2024

Vantive and Extraneal are trademarks of Vantive Health LLC or its affiliates.

**Vantive**