

## Country-Specific Glucose Monitor List

**Country Name:** *Philippines*

### **Important Information**

This is a non-comprehensive list, current as of **November 4, 2022**. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with Icodextrin and Maltose, Baxter does NOT recommend that EXTRANEAL patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitors manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with Maltose and Icodextrin to Baxter's recommended limits: 278 mg/dL (Maltose) and 1094 mg/dL (Icodextrin). The manufacturers certified that their "green" monitors mentioned below showed no interference of blood glucose readings. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, Literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) Peritoneal Dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains Icodextrin. Maltose, a metabolite of Icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by Icodextrin or Maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. ONLY glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution. Contact the manufacturer to verify that the test strips and monitor are glucose specific. The list provides the contact information of the more common, major brand manufacturers.

## COUNTRY-SPECIFIC GLUCOSE MONITORS

Updated November 4, 2022

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
FreeStyle Freedom	Yes	GDH-FAD	<b>Abbott Diabetes Care</b> <a href="http://www.abbottdiabetescare.com">www.abbottdiabetescare.com</a>  <a href="https://www.diabetescare.abbott/worldwide-locations/contact.philippines.html">https://www.diabetescare.abbott/worldwide-locations/contact.philippines.html</a>  <b>Phone:</b> +63 (02)-8702-8577
FreeStyle Freedom Lite	Yes	GDH-FAD	
FreeStyle Lite	Yes	GDH-FAD	
FreeStyle Libre <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre Pro <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre 2 <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre 3 <sup>1</sup>	Not Recommended	GO	
FreeStyle Optium Neo	Yes	GDH-NAD	
FreeStyle Optium Neo H	Yes	GDH-NAD	
FreeStyle Papillon Vision	Yes	GDH-FAD	
FreeStyle Precision Neo	Yes	GDH-NAD	
FreeStyle Precision Pro	Yes	GDH-NAD	
Optium Xido Neo	Yes	GDH-NAD	
Precision Xceed Pro	Yes	GDH-NAD	
Assure Platinum	Yes	GO	<b>ARKRAY Co. Ltd., Inc.</b> <a href="https://www.arkray.ph/english/">https://www.arkray.ph/english/</a>  <b>Phone:</b> +63-(02)-8843-4731  <b>Arkray Industry, Inc.</b> <a href="https://www.arkray.ph/english/index.html">https://www.arkray.ph/english/index.html</a>  <b>Phone:</b> +63-(43)-405-6450 +63-(43)-405-6449
GLUCOCARD 01	Yes	GO	
GLUCOCARD 01-mini	Yes	GO	
GLUCOCARD 01-mini plus	Yes	GO	
GLUCOCARD MX	Yes	GDH-FAD	
GLUCOCARD S	Yes	GDH-FAD	
GLUCOCARD SM	Yes	GDH-FAD	
GLUCOCARD Vital	Yes	GO	
GLUCOCARD X-meter	Yes	GDH-FAD	
GLUCOCARD X-mini	Yes	GDH-FAD	
GLUCOCARD X-mini plus	Yes	GDH-FAD	
GLUCOCARD Σ	Yes	GO	
GLUCOCARD Σ-mini	Yes	GO	
ReliOn Confirm	Yes	GO	
ReliOn micro	Yes	GO	
ReliOn Prime	Yes	GO	
GLUCOCARD S onyx	Yes	GDH-FAD	
PocketChem BG	Yes	GDH-FAD	
GLUCOCARD W	Yes	GO	
GLUCOCARD PlusCare	Yes	GDH-FAD	
GLUCOCARD Prime	Yes	GDH-FAD	
GlucoCard G+	Yes	GO	
GlucoCard G+ Link	Yes	GO	
GlucoCard Σ+	Yes	GO	
GlucoCard Σ+ Link	Yes	GO	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
<b>Breeze 2</b> <b>Contour</b> <b>Contour Link</b> <b>Contour Next</b> <b>Contour Next EZ</b> <b>Contour Next Link</b> <b>Contour Next Link 2.4</b> <b>Contour Next USB</b> <b>Contour Next One</b> <b>Contour Plus</b> <b>Contour Plus Link 2.4</b> <b>Contour Plus One</b> <b>Contour TS</b> <b>Contour USB</b> <b>Contour XT</b> <b>Contour Next (Connected)</b> <b>Contour Care</b> <b>Contour Plus Blue</b> <b>Contour Plus Elite</b> <b>Contour Next GEN</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Ascensia Diabetes Care</b> (formerly Bayer Healthcare) <a href="https://www.diabetes.ascensia.ph/">https://www.diabetes.ascensia.ph/</a>  <b>Phone:</b> +63 (02) 8982-7000  <b>Email:</b> <a href="mailto:customerservice.my@ascensia.com">customerservice.my@ascensia.com</a>
<b>OneTouch Select</b> <b>OneTouch Select Mini</b> <b>OneTouch SelectPlus</b> <b>OneTouch SelectPlus Flex</b> <b>OneTouch SelectPlus Simple</b> <b>OneTouch Select Simple</b> <b>OneTouch Ultra<sup>A</sup></b> <b>OneTouch Ultra 2</b> <b>OneTouch UltraEasy</b> <b>OneTouch UltraMini</b> <b>OneTouch UltraPlus Flex</b> <b>OneTouch UltraPlus Reflect</b> <b>OneTouch UltraVue</b> <b>OneTouch Verio</b> <b>OneTouch VerioFlex</b> <b>OneTouch VerioIQ</b> <b>OneTouch VerioPro+</b> <b>OneTouch VerioVue</b> <b>OneTouch Verio Reflect</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Lifescan, Inc.</b> <a href="http://www.lifescan.com">www.lifescan.com</a>  <b>Phone:</b> +63 (02) 7368-8000 ext. 2 +63 917 878 2218 +6318 001 888 7954

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Nova Max Plus	Yes	GO	Nova Biomedical
Nova Max Link	Yes	GO	<a href="http://www.novabiomedical.com">www.novabiomedical.com</a>
StatStrip Hospital Glucose Meter	Yes	GO	
StatStrip Hospital Glucose and Ketone Monitoring System	Yes	GO	Phone: (63) 0995-734-7835
StatStrip Xpress Glucose Meter	Yes	GO	
StatStrip Xpress Glucose and Ketone Monitoring System	Yes	GO	
StatStrip Xpress2 Glucose and Ketone Monitoring System	Yes	GO	
Nova Pro Glucose/Ketone Meter	Yes	GDH-FAD	
Accu-Chek Active	Yes	Mut Q-GDH	Roche (Philippines) Inc. Diagnostics Division
Accu-Chek Aviva	Yes	Mut Q-GDH	<a href="http://www.accu-chek.com">www.accu-chek.com</a>
Accu-Chek Aviva Combo	Yes	Mut Q-GDH	
Accu-Chek Aviva Connect	Yes	Mut Q-GDH	<a href="https://diagnostics.roche.com/ph/en/about/Contact-Us.html">https://diagnostics.roche.com/ph/en/about/Contact-Us.html</a>
Accu-Chek Aviva Expert	Yes	Mut Q-GDH	
Accu-Chek Aviva Insight	Yes	Mut Q-GDH	<a href="mailto:manila.diagnostics@roche.com">manila.diagnostics@roche.com</a>
Accu-Chek Aviva Plus <sup>2</sup>	Yes	Mut Q-GDH	
Accu-Chek Aviva Solo	Yes	Mut Q-GDH	
Accu-Chek Guide	Yes	GDH-FAD	Phone: +63 (02) 7718-7777
Accu-Chek Guide Me	Yes	GDH-FAD	
Accu-Chek Inform II	Yes	Mut Q-GDH	Service Excellence Center
Accu-Chek Instant	Yes	GDH-FAD	Toll-Free Hotline: #76243
Accu-Chek Instant S	Yes	GDH-FAD	(#ROCHE)
Accu-Chek Nano <sup>2</sup>	Yes	Mut Q-GDH	
Accu-Chek Nano SmartView <sup>2</sup>	Yes	Mut Q-GDH	
Accu-Chek Mobile	Yes	Mut Q-GDH	
Accu-Chek Performa	Yes	Mut Q-GDH	
Accu-Chek Performa Combo	Yes	Mut Q-GDH	
Accu-Chek Performa Connect	Yes	Mut Q-GDH	
Accu-Chek Performa Insight	Yes	Mut Q-GDH	
Accu-Chek Performa Solo	Yes	Mut Q-GDH	
Accu-Chek Guide Solo	Yes	GDH-FAD	
TRUE METRIX	Yes	GDH-FAD	Trividia Health
TRUE METRIX GO	Yes	GDH-FAD	<a href="http://www.trividiahealth.com">www.trividiahealth.com</a>
TRUE METRIX Air	Yes	GDH-FAD	
TRUEresult <sup>3</sup>	Not Recommended	GDH-FAD	Phone: 1-800-803-6025
TRUEresult Twist 2 <sup>3</sup>	Not Recommended	GDH-FAD	
TRUEyou <sup>3</sup>	Not Recommended	GDH-FAD	
TRUEtrack <sup>3</sup>	Not Recommended	GO	
TRUEresult	No	GDH-PQQ	

<sup>1</sup> These Abbott monitors (together with the Dexcom G4 Platinum, G5 Mobile, G6 and the Medtronic iPro2) are the leading Continuous Glucose Monitor Systems (CGM) and are **not recommended** by their manufacturers for use into dialysis patients, and despite having a GO based chemistry currently there is not definitive scientific evidence for their Extraneal compatibility. Consult manufacturer for additional information.

<sup>2</sup> The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY and use test strips branded as ACCU-CHEK SmartView and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**Compatible**) strips. Consult manufacturer for additional information.

<sup>3</sup> TRUEresult (GDH-FAD), TRUEresult Twist 2, TRUEyou, True track monitors were not tested with the Baxter recommended limits for Icodextrin and Maltose. So, these monitors were not recommended to use for dialysis patients.

#### **Test Type\***

**GO** = Glucose Oxidase

**GDH-PQQ** = Glucose Dehydrogenase with Pyrroloquinolinequinone (note: **GDO**, Glucose-dye-oxidoreductase, is an **incompatible** PQQ-based method)

**GDH-NAD** = Glucose dehydrogenase with nicotinamide-adenine dinucleotide

**GDH-FAD** = Glucose dehydrogenase with flavin-adenine dinucleotide

**Mut Q-GDH** = Glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

#### **References:**

<sup>A</sup> Baxter report Interim 1, 33541 Determination of potential interference of Icodextrin and its metabolites on human blood glucose measurement using chosen glucometers.

Please see full prescribing information for EXTRANEAL (Icodextrin) PD solution.

Baxter and EXTRANEAL are trademarks of Baxter International Inc.

Other trademarks or products appearing herein are the property of their respective owners.