

Country-Specific Glucose Monitor List

Country Name: *United States*

Important Information

This is a non-comprehensive list, current as of **December, 2021**. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with icodextrin and maltose, Baxter does NOT recommend that EXTRANEAL patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitor manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with maltose and icodextrin to Baxter's recommended limits: 278 mg/dL (Maltose) and 1094 mg/dL (Icodextrin). The manufacturers certified that their "green" monitors mentioned below showed no interference of blood glucose readings. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, Literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) Peritoneal Dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains Icodextrin. Maltose, a metabolite of Icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely-elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by Icodextrin or Maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. **ONLY** glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution. Contact the manufacturer to verify that the test strips and monitor are glucose-specific. The list provides the contact information of the more common, major brand manufacturers.

GLUCOSE MONITORS

Updated December 2021

| Glucose Monitor Brand | Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific) | Test Type* | Manufacturer |
|--|--|----------------|--|
| FreeStyle Freedom | Yes | GDH-FAD | Abbott Diabetes Care www.abbottdiabetescare.com Phone: 888-522-5226 |
| FreeStyle Freedom Lite | Yes | GDH-FAD | |
| FreeStyle Lite | Yes | GDH-FAD | |
| FreeStyle Libre¹ | Not Recommended | GO | |
| FreeStyle Libre Pro¹ | Not Recommended | GO | |
| FreeStyle Libre 2¹ | Not Recommended | GO | |
| FreeStyle Libre 3¹ | Not Recommended | GO | |
| FreeStyle Optium Neo | Yes | GDH-NAD | |
| FreeStyle Optium Neo H | Yes | GDH-NAD | |
| FreeStyle Papillon Vision | Yes | GDH-FAD | |
| FreeStyle Precision Neo | Yes | GDH-NAD | |
| FreeStyle Precision Pro | Yes | GDH-NAD | |
| Optium Xido Neo | Yes | GDH-NAD | |
| Precision Xceed Pro | Yes | GDH-NAD | |
| Assure Platinum | Yes | GO | Arkray, Inc. www.arkray.co.jp Phone: 800-818-8877 |
| GLUCOCARD 01 | Yes | GO | |
| GLUCOCARD 01-mini | Yes | GO | |
| GLUCOCARD 01-mini plus | Yes | GO | |
| GLUCOCARD MX | Yes | GDH-FAD | |
| GLUCOCARD S | Yes | GDH-FAD | |
| GLUCOCARD SM | Yes | GDH-FAD | |
| GLUCOCARD Vital | Yes | GO | |
| GLUCOCARD X-meter² | Yes | GDH-FAD | |
| GLUCOCARD X-mini² | Yes | GDH-FAD | |
| GLUCOCARD X-mini plus² | Yes | GDH-FAD | |
| GLUCOCARD Σ | Yes | GO | |
| GLUCOCARD Σ-mini | Yes | GO | |
| ReliOn Confirm | Yes | GO | |
| ReliOn micro | Yes | GO | |
| ReliOn Prime | Yes | GO | |
| GLUCOCARD S onyx | Yes | GDH-FAD | |
| PocketChem BG | Yes | GDH-FAD | |
| GLUCOCARD W | Yes | GO | |
| GLUCOCARD PlusCare | Yes | GDH-FAD | |
| GLUCOCARD Prime | Yes | GDH-FAD | |
| GLUCOCARD G+ | Yes | GO | |
| GLUCOCARD G+ Link | Yes | GO | |

| Glucose Monitor Brand | Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific) | Test Type* | Manufacturer |
|-----------------------------------|---|----------------|--|
| Breeze 2 | Yes | GO | Ascensia Diabetes Care (formerly Bayer Healthcare) www.ascensia.com Phone: 800-348-8100 |
| Contour | Yes | GDH-FAD | |
| Contour Link | Yes | GDH-FAD | |
| Contour Next | Yes | GDH-FAD | |
| Contour Next EZ | Yes | GDH-FAD | |
| Contour Next Link | Yes | GDH-FAD | |
| Contour Next Link 2.4 | Yes | GDH-FAD | |
| Contour Next USB | Yes | GDH-FAD | |
| Contour Next One | Yes | GDH-FAD | |
| Contour Plus | Yes | GDH-FAD | |
| Contour Plus Link 2.4 | Yes | GDH-FAD | |
| Contour Plus One | Yes | GDH-FAD | |
| Contour TS | Yes | GDH-FAD | |
| Contour USB | Yes | GDH-FAD | |
| Contour XT | Yes | GDH-FAD | |
| Contour Next (Connected) | Yes | GDH-FAD | |
| Contour Care | Yes | GDH-FAD | |
| OneTouch Select | Yes | GO | Lifescan, Inc. www.lifescan.com Phone: 800-227-8862 |
| OneTouch Select Mini | Yes | GO | |
| OneTouch SelectPlus | Yes | GO | |
| OneTouch SelectPlus Flex | Yes | GO | |
| OneTouch SelectPlus Simple | Yes | GO | |
| OneTouch Select Simple | Yes | GO | |
| OneTouch Ultra^A | Yes | GO | |
| OneTouch Ultra 2 | Yes | GO | |
| OneTouch UltraEasy | Yes | GO | |
| OneTouch UltraLink | Yes | GO | |
| OneTouch UltraMini | Yes | GO | |
| OneTouch UltraPlus Flex | Yes | GDH-FAD | |
| OneTouch UltraPlus Reflect | Yes | GDH-FAD | |
| OneTouch UltraSmart | Yes | GO | |
| OneTouch UltraVue | Yes | GO | |
| OneTouch Verio | Yes | GDH-FAD | |
| OneTouch VerioFlex | Yes | GDH-FAD | |
| OneTouch VerioIQ | Yes | GDH-FAD | |
| OneTouch VerioPro | Yes | GDH-FAD | |
| OneTouch VerioPro+ | Yes | GDH-FAD | |
| OneTouch VerioSync | Yes | GDH-FAD | |
| OneTouch VerioVue | Yes | GDH-FAD | |
| OneTouch Verio Reflect | Yes | GDH-FAD | |

| Glucose Monitor Brand | Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific) | Test Type* | Manufacturer |
|---|---|------------|---|
| Nova Max Plus | Yes | GO | Nova Biomedical www.novabiomedical.com Phone : 781-894-0800 |
| Nova Max Link | Yes | GO | |
| StatStrip Hospital Glucose Meter | Yes | GO | |
| StatStrip Hospital Glucose and Ketone Monitoring System | Yes | GO | |
| StatStrip Xpress Glucose Meter | Yes | GO | |
| StatStrip Xpress Glucose and Ketone Monitoring System | Yes | GO | |
| StatStrip Xpress2 Glucose and Ketone Monitoring System | Yes | GO | |
| Nova Pro Glucose/Ketone Meter | Yes | GDH-FAD | |
| Accu-Chek Active | Yes | Mut Q-GDH | |
| Accu-Chek Aviva | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Combo | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Connect | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Expert | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Insight | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Plus ³ | Yes | Mut Q-GDH | |
| Accu-Chek Aviva Solo | Yes | Mut Q-GDH | |
| Accu-Chek Guide | Yes | GDH-FAD | |
| Accu-Chek Guide Me | Yes | GDH-FAD | |
| Accu-Chek Inform II | Yes | Mut Q-GDH | |
| Accu-Chek Instant | Yes | GDH-FAD | |
| Accu-Chek Instant S | Yes | GDH-FAD | |
| Accu-Chek Nano ³ | Yes | Mut Q-GDH | |
| Accu-Chek Nano SmartView ³ | Yes | Mut Q-GDH | |
| Accu-Chek Mobile | Yes | Mut Q-GDH | |
| Accu-Chek Performa | Yes | Mut Q-GDH | |
| Accu-Chek Performa Combo | Yes | Mut Q-GDH | |
| Accu-Chek Performa Connect | Yes | Mut Q-GDH | |
| Accu-Chek Performa Insight | Yes | Mut Q-GDH | |
| Accu-Chek Performa Solo | Yes | Mut Q-GDH | |
| Accu-Chek Guide Insight | Yes | GDH-FAD | |
| TRUE METRIX | Yes | GDH-FAD | Trividia Health www.trividiahealth.com Phone: 800-803-6025 |
| TRUE METRIX GO | Yes | GDH-FAD | |
| TRUE METRIX Air | Yes | GDH-FAD | |
| TRUEresult ⁴ | Not Recommended | GDH-FAD | |
| TRUEresult Twist 2 ⁴ | Not Recommended | GDH-FAD | |
| TRUEyou ⁴ | Not Recommended | GDH-FAD | |
| TRUEtrack ⁴ | Not Recommended | GO | |
| TRUEresult | No | GDH-PQQ | |

| Glucose Monitor Brand | Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific) | Test Type* | Manufacturer |
|--|---|---------------------|--|
| Jazz Jazz Wireless 2 Presto iBGStar One Drop | No compatibility certification form completed by AgaMatrix, Inc. in 2021, contact manufacturer to confirm compatibility | | AgaMatrix, Inc. www.agamatrix.com Phone: 866-906-4197 |
| Prodigy AutoCode Prodigy Voice Prodigy Pocket | Yes Yes Yes | GO GO GO | Prodigy Diabetes Care www.prodigymeter.com Phone: 800-243-2636 |
| Embrace Pro Embrace TALK Embrace | Yes Yes Yes | GDH-FAD GO GO | Omnis Health www.omnishealth.com Phone: 877-979-5454 |

¹ These Abbott monitors (together with the Dexcom G4 Platinum, G5 Mobile, G6 and the Medtronic iPro2) are the leading Continuous Glucose Monitor Systems (CGM) and are **not recommended** by their manufacturers for use into dialysis patients, and despite having a GO based chemistry currently there is not definitive scientific evidence for their Extraneal compatibility. Consult manufacturer for additional information.

² These Arkray monitors/test strips have switched from an **incompatible** GDH-FAD based chemistry to a **compatible** GDH-FAD based chemistry. Consult manufacturer for additional information.

³ The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY and use test strips branded as ACCU-CHEK SmartView and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**Compatible**) strips. Consult manufacturer for additional information.

⁴ TRUEresult (GDH-FAD), TRUEresult Twist 2, TRUEyou, True track monitors were not tested with the Baxter recommended limits for Icodextrin and Maltose. So, these monitors were not recommended to use for dialysis patients.

Test Type*

GO = glucose oxidase

GDH-PQQ = glucose dehydrogenase with pyrroloquinolinequinone (note: **GDO**, glucose-dye-oxidoreductase, is an **incompatible** PQQ-based method)

GDH-NAD = glucose dehydrogenase with nicotinamide-adenine dinucleotide

GDH-FAD = glucose dehydrogenase with flavin-adenine dinucleotide

Mut Q-GDH = glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

References:

^A Baxter report Interim 1, 33541 Determination of potential interference of icodextrin and its metabolites on human blood glucose measurement using chosen glucometers.

Please see full prescribing information for EXTRANEAL (icodextrin) PD solution.

Baxter and EXTRANEAL are trademarks of Baxter International Inc.

Other trademarks or products appearing herein are the property of their respective owners.