

WARNING



Baxter

Potential for Incorrect Blood Glucose Reading
(regardless of cardholder's diabetic status)



WARNING

Potential for Incorrect
Blood Glucose Reading

Baxter

Patient Name

is using **EXTRANEAL (icodextrin) Peritoneal Dialysis Solution**

Use blood glucose readings obtained from laboratory-based methods or compatible glucometers and test strips. Use of incompatible glucometers and test strips may result in falsely elevated blood glucose readings for this patient.

Name of Nephrologist

Phone Number

Name of PD Nurse/Center

Phone Number

Please visit www.glucosafety.com or contact 1-888-736-2543 (Option 1)

Baxter and Extraneal are trademarks of Baxter International Inc.

07-27-74-132 08/15

ATTENTION MEDICAL PROFESSIONAL

Patient is using EXTRANEAL (icodextrin)
Peritoneal Dialysis Solution

Please visit www.glucosafety.com
or contact 1-888-736-2543 (Option 1)