

# WARNING

## Potential for Incorrect Glucose Reading

Patient Name

Is using **EXTRANEAL** (icodextrin) peritoneal dialysis solution

**ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.**

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (icodextrin) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events.

Visit [www.glucosafety.com](http://www.glucosafety.com) for additional information, including a glucose monitor compatibility

SG-RC21-230011

Date of Preparation Sept 2023

Baxter and **EXTRANEAL** are trademarks of Baxter International Inc.

**Baxter**